

Remove the collimator and check in the tubes for spider webs or obstructions and gently clean inside them using a soft bottle brush taking care not to dislodge the baffles inside the collimator (*when removing the collimator, observe the orientation of the collimator for reinstallation*). Also inspect the lenses for any moisture, dust or debris but please do not touch them. The lenses should never be wiped off as this will affect the calibration. If something is on the lenses please send me a picture to see if any further action is required.

Gently clean inside the collimator using a soft bottle brush taking care not to dislodge the baffles inside

