

TRACEABILITY CHAIN OF THE WMO AOD REFERENCE AND GAW-PFR NETWORK

**Natalia Kouremeti¹, Stelios Kazadzis¹, Julian Gröbner¹,
Saulius Nevas², Kerstin Schwind², Philipp Schneider²**

¹Physikalisch-Meteorologisches Observatorium Davos, World Radiation Center, PMOD/WRC

²Physikalisch-Technische Bundesanstalt (PTB), Braunschweig, Germany

Precision Filter Radiometers (PFR) description


- Filter radiometer with 4 channels in a grid
 - Si-photodiodes
 - Dielectric interference filters manufactured by the ion-assisted deposition technique
- Centroid wavelength based on WMO recommendation

Standard (N) 368 nm, 412 nm, 500 nm, 862 nm

Band pass at FWHM : 3 nm to 5 nm

- Sealed tube – purged with N₂ to 2bar
 - Temperature stabilized photodiodes (20°C either 25°C)
 - Well defined reference plane – precision aperture
 - Two amplification stages
-
- Sun sensor for monitoring alignment to Sun
 - FOV : 1.2° plateau, 0.7° slope angle, inhomogeneity in plateau < 0.5%
-
- Internal PFR pressure monitor
 - Ambient atmospheric pressure

**Direct Solar irradiance
measurements
1 min time resolution**



Data Acquisition Systems

- Campbell Scientific (CR10X, CR1000, CR1000x)
- SACRAM 22-bit data acquisition system specifically designed for the PFR.

WMO AOD reference – GAW-PFR_TRIAD at PMOD/WRC, Davos

Service & Calibration

Service & Calibration




GAW-PFR-Triad = 9 PFR instruments

- Almost identical spectral characteristics (wavelength, bandwidth)
- Well maintained
- The scale is realized by comparison to Langley standards.
- The stability and precision are assessed annually.
- Pool of 9 traveling standards for deployment at the Langley calibration sites.



Izaña, Tenerife (ES)




Mauna-Loa, Hawaii (US)



GAW_PFR-Triad – WMO AOD reference



GAW-PFR-Triad calibration stability better than 1%




Traceability of GAW-PFR-Triad and GAW-PFR network instruments

International vocabulary of metrology (VIM)

Metrological Traceability:

Property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty

- ✓ Unbroken chain of comparisons
- ✓ Uncertainty of measurement
- ✓ Reference to SI units
- ✓ Calibration intervals
- Documentation for each step in the chain
- Prove competence of realizing the chain



19ENV04 MAPP
Metrology for aerosol
optical properties




Traceability of the GAW-PFR
reference Precision
FilterRadiometers to the SI

Langley Calibration - Uncertainty

$$\ln(I_\lambda) + \sum_i \tau_{att(i)} m_{att(i)} = \ln(I_\lambda^0) - AOD_\lambda m$$

Component	Description	Type	Value
Signal Noise Atmospheric	Standard deviation of the 10 sequential measurements	A	
Dark Noise	Standard deviation of the day	A	
Pressure	Uncertainty of pressure measurement	B	3mbar
Total column of ozone	Daily value of ozone measurement (OMI)	B	2%
Ozone airmass	Height of ozone layer	B	3km
Airmass			
Equation	Comparison of 2 different algorithms	B	<0




Wavelength(nm)	862	500	412	368
Langley regression expanded uncertainty (% ,k=2)	0.003	0.019	0.038	0.060

PFR signal uncertainty

example for 500 nm

Component	Description	Type	Symbol	Sensitivity Coefficient	DF	units	Value
Signal combined relative uncertainty							0.24
Reference PFR uncertainty							6
Combined relative uncertainty							0.273
Measurement repeatability	Standard deviation of the 10 sequential measurements	A	1	9	%	0.010	
Dark signal noise	Standard deviation of the day / period				%	0.05	
PFR-DAQ	Standard deviation of the reference 2500 mV over the calibration period	B		1	%	0.008	
Electronic Noise							
FOV	0.2 % within ±10 arcmin of the center of the plateau	B		1	%	0.200	
DAQ precision	0.1% signal range 1-5000 mV	B		1	%	0.010	
Non-Linearity PFR-Gain	Gain uncertainty : estimated to 0.1% at PTR	B		1	%	0.100	




Calibration of reference PFRs against 3-Langley transfer PFRs

Component	Description	Type	Symbol	Sensitivity Coefficient	DF	units	Value
GAW-PFR network & customer PFRs							
Corrections for wavelength mismatch (368 nm, 412 nm)							
Deviation between uncorrected V_0 and							
Signal Ratio uncertainty	<ul style="list-style-type: none"> Radiative transfer model correction and airmass extrapolation of the ratios. 						
Daily mean ratio	standard deviation of the signals' ratios acquired during one day	A			N_i	%	0.100
				1	N_j	%	0.472
Combined relative uncertainty		A			N_i	%	0.273


GAW-PFR TRIAD scale transfer to GAW-PFR network




Provide Calibration traceable to WMO AOD reference.




Izaña, Spain



OHP, France




Valladolid, Spain




- PFR AOD uncertainty
 - calibration, signal, pressure, airmass, O₃, NO₂

Work in progress


- Revise uncertainties of PFR for the extrapolated wavelength 340 nm (Monte Carlo approach on linear fit - add 2nd order fit)
- Optimize PFR operation in OHP, VLD




SI-traceable Solar irradiance



Scanned spectral range: 320 nm to 1020 nm




Top-of-Atmosphere High Resolution Solar Irradiance Spectrum




Tunable Lasers In Metrology (TULIP)

- fully automated
- based on
- homodyne
- wave
- wave


Calibration uncertainty < 0.3%



SI-traceable Solar irradiance



	Change in responsivity (%)				Change in centroid wavelength (pm)			
	863.0	500.8	412.0	367.6	863.0	500.8	412.0	367.6
2018 (-3 years)	0.27	0.40	0.81	0.80	-58	10	5	0
2021	0	0	0	0	0	0	0	0
2024 (+3 years)	-0.28	-0.34	-0.92	-1.00	0	0	2	-2



SI- and Langley based calibrations produce equivalent AOD within the WMO requested uncertainty.

Other check points

- 2021, FRC-V (GAW Report 208)
- 2022, MAPP Izaña campaign (Gröbner, et.al,2023)

- Working on establishing Metrological Traceability for the GAW-PFR references and field instrument
- SI-traceable direct solar irradiance and AOD retrievals give equivalent result to the Langley calibration
- Disseminate the traceability through AOD comparison at ACTRIS/CARS sites, and exploring possibilities of SI-calibration transfer



Thank you for your attention !





AERONET Science and Application Exchange 2024, September 17-19, 2024

~20 years


Mauna Loa, USA





Izaña, Spain



Data Quality Assurance - Sources of Highest uncertainties





ch1: Cntr=-0.07°, Plateau=1.2°, U=0.20%(>0.99 of max) FWHM=2.41 [-1.25, 1.16]




$$LR = cf \left(\frac{((Q_1 + Q_3) - (Q_2 + Q_4))}{\sum_{i=1}^4 Q_i} - LRO \right)$$

$$UD = cf \left(\frac{((Q_1 + Q_2) - (Q_3 + Q_4))}{\sum_{i=1}^4 Q_i} - UDO \right)$$



GAW-PFR TRIAD scale maintenance



$$0.9 \leq AE \leq 2.2 \text{ and } AOD(500 \text{ nm}) < 0.5$$

The selected synchronized PFR -TS AOD at AOD data at the wavelength of the DUT are subtracted from the DUT AOD. A control of differences with respect to errors in the quality control of both PFR-TS and DUT is applied, excluding extreme values. Days with more than 10 synchronous measurements are selected. The atmospheric conditions filter with respect the aerosol load and uncertainty of extra\interpolation is applied on the differences :




Figure : Example of combined uncertainty introduced to extra/interpolated AOD to 340nm,380 nm, 440 nm, 675 nm 1020 nm, due to the extrapolation method and uncertainty the measured atmospheric transmittance by the PFR at 368 412 nm, 500 nm and 862 nm with a combined expanded relative irradiance uncertainty of 0.32%, 0.32% , 0.30% 0.2% respectively (AOD at 500 nm = 0.01).

GAW-PFR TRIAD scale maintenance

